



Cello

PC & iPad CONTROLLED DIAGNOSTIC AUDIOMETER

CLASSIFICATION

EN 60645-1 / ANSI S3.6: Type 2
 EN 60645-2 / ANSI S3.6: Type A or A-E
 EN 60645-4 (with HF option)

AVAILABLE SIGNALS

Stimulus: pure tone, warble tone
 2 external inputs and USB audio for speech audiometry
 MIC input for live speech audiometry
 Masking: NBN, WN, SN

SIGNALS SPECIFICATION

Attenuator step: 1 and 5 dB
 Presentation: Continuous, Pulsed (0.5, 1 and 2 Hz)
 Warble: 5 Hz sin wave modulating signal

AVAILABLE OUTPUTS AND TRANSDUCERS

AC: Headphones (TDH-39 / DD45)
 HF headphones (HDA-300)
 Insert earphones (ER-3A / ER-5A)
 BC: B-71 bone vibrator
 Free field
 Insert masking earphone: IME-100

AVAILABLE TESTS

- Pure Tone audiometry (HL and UCL)
- Auto threshold (modified Hughson-Westlake)
- HF audiometry: from 8 to 20 kHz (optional)
- Speech audiometry (SRT, WRS, MCL, and UCL)
- SISI test
- 2 independent channels Master Hearing Aid
- QuickSIN® test (optional)
- Visual Reinforcement Audiometry (VRA) and Conditioned Play Audiometry (CPA) test (optional) with video reinforces

PURE TONE: FREQUENCIES AND MAXIMUM LEVELS (dB HL)

Freq. (Hz)	AC TDH-39 / DD45	AC HDA-300	AC ER-3A	AC ER-5A	BC	FF (*)
125	80	85	90	90	-	75
250	100	100	105	100	45	85
500	110	120	110	110	65	95
750	115	120	115	120	70	95
1.000	120	120	120	120	75	95
1.500	120	110	120	120	80	95
2.000	120	110	120	115	80	95
3.000	120	110	120	115	75	95
4.000	120	110	110	110	75	95
6.000	105	100	100	100	55	90
8.000	95	100	90	90	50	85
9.000	-	90	-	-	-	80
10.000	-	90	-	-	-	80

11.200	-	90	-	-	-	80
12.500	-	80	-	-	-	80
14.000	-	70	-	-	-	80
16.000	-	60	-	-	-	50
18.000	-	100 dB SPL	-	-	-	-
20.000	-	90 dB SPL	-	-	-	-

(*) The values refer to "normal" range; add 10 dB to each value in case of "extended range" option selected

SPEECH AUDIOMETRY: MAXIMUM LEVELS (dB HL)

AC TDH-39 / DD45	AC HDA-300	AC ER-3A	AC ER-5A	BC	FF
100	100	100	100	55	Normal: 75 Extended: 85

PATIENT – OPERATOR COMMUNICATION

Talk over: external microphone included
 Talk back: through monitor headset (included); patient microphone included
 Patient response trigger

MONITOR SIGNAL

Both channels monitored through the monitor headset (included)

CALIBRATION

Validity: 12 months
 Separated calibration tables for each type of transducer (AC, AC-INS, AC-HF, INS-MASK, BC, FF)
 All the parameters stored in the device through a dedicated software

COMPUTER INTERFACE

Connection: USB (driverless)
 Compatible software: - Inventis Daisy with Maestro module
 - Noah with "Maestro for Noah" module

INTERFACING WITH THE iPad

Through Bluetooth 4 connection (needs *Aero* option).
 Requires iPad 3 or successive
 Dedicated app freely downloadable from the Apple App Store
 Only Pure Tone and Speech tests available

POWER SUPPLY

Maximum consumption: 10 Watts
 Power supply: 6V, 2A cont., through an external medical grade 100-240 Vac 50/60 Hz power supply

MECHANICS

Size (WxDxH): 30.8 x 21.7 x 6.7 cm / 12.1 x 8.5 x 2.5 in
 Weight: 1.3 kg / 2.9 lbs

APPLICABLE STANDARDS

Pure tone audiometry: EN 60645-1, Type 2
 Speech audiometry: EN 60645-2, Type A or A-E (depending on the



equalization filter status)
 Calibration: EN ISO 389-1 (TDH 39), EN ISO 389-2 (ER-3A and ER-5A), EN ISO 389-3 (B71), EN ISO 389-7 (FF), data from the manufacturer (DD45 and HDA-300 headphones)
 Electrical safety: EN 60601-1
 EMC: EN 60601-1-2

CE CERTIFICATE

93/42/EEC classification : Class IIa
 Classification rule (Annex IX, 93/42/EEC): 10
 Notified body: TÜV SÜD Product Service GmbH (0123)
 Number of CE certificate: G1 12 09 65346 006

PRODUCT CODES

10558: Cello – Computer controlled diagnostic audiometer
 10559: Cello Aero – Computer /iPad controlled diagnostic audiometer

INCLUDED PARTS

- TDH-39 supra-aural headphones
- B71 bone vibrator
- Patient response switch
- Monitor headset with boom microphone
- Clip-on microphone for patient-to-operator communication
- Medical grade power supply
- USB connection cable
- User manual
- Inventis Software Suite CD
- Noah Maestro module
- Daisy software and Daisy Maestro module
- Four magnetic covers

OPTIONAL ACCESSORIES (with order code)

- 10166: ER-3A insert earphones
- 10172: ER-5A insert earphones
- 10177: IME-100 insert masking earphone
- 10561: DD45 complete headset
- 10179: Amplivox Audiocups noise excluding enclosures for TDH-39 and DD45 headphones
- 10181: Desktop microphone for live speech tests
- 10180: Cable set for sound booths with jack sockets
- 10557: Cable set for sound booths with RCA sockets
- 10266: One active speaker FBT J 5A (230V model)
- 10544: One active speaker FBT J 5A (120V model)
- 10560: High-Frequency extension package for Cello
- 10533: QuickSIN® test license
- 10562: Video-VRA extension package for Cello. Includes the license code and the desktop switch.

Cello is developed by:

INVENTIS s.r.l.
 CORSO STATI UNITI, 1/3
 35127 PADOVA – ITALY
 PHONE: +39.049.8962 844
 FAX: +39.049.8966 343
 info@inventis.it
 www.inventis.it

Follow us on Twitter
<http://twitter.com/inventissrl>



The Inventis Quality System complies with ISO 9001 and ISO 13485 standards.

Inventis® is a registered trademark of Inventis s.r.l.